



May 28th, 2015

Product Quality Certificate

Product code: IT26-1YL

Product description: Unique point Integrator Indicator for Steam sterilization processes. Class 5 (EN ISO 11140-1:2009).

Batch number: IT261YL270515

Manufacture date: 05.2015

Expiration date: 05.2020

Transition color: Turns to Black/Vira a Negro.

Process specification:

SV121°C = 19.2 min

SV128°C = 8.9 min

SV134°C = 4.4 min

Parameters determined at the time of manufacturing in accordance with USP and ISO 11140-1:2014 standards. The values shown in this certificate are reproducible only under the conditions in which they were established.

CONCLUSION: Product batch **IT261YL270515** meets the specifications under USP and ISO 11140-1:2014 regulation.



TERRAGENE S.A.

Lic. Adrián Rovetto
DIRECTOR TÉCNICO