

Automatic System for Quality & Traceability Control of Washing & Sterilization processes

Trazanto® Compatible Indicators



Washing



CDWA3

For Washing Efficiency tests

CDWA4

For Washing Efficiency tests



LUMENIA

For Washing Efficiency control in flexible endoscopes' hollow internal channels

Sterilization



CD29

For Steam



CD40

For Hydrogen Peroxide



CD42

For Hydrogen Peroxide



CD50

For Formaldehyde



**PCD20-2 | PCD20-C
PCD220-2 | PCD220-C
PCD222-2 | PCD222-C
PCD224-2 | PCD224-C
PCD225-2 | PCD224-C**

For Steam processes



PCD26-2 | PCD26-C

For Steam processes



BD125X/1

Bowie-Dick Test Pack for 3.5 min at 134 °C



BD125X/2

Bowie-Dick Test Pack for 4 min at 132 °C & for 3.5 min at 134 °C



IT26-1YS

For Steam processes between 121-135 °C



IT26-SBL

Unique point Steam Type 5 CI For Steam processes between 121-135 °C



IT26-SAD

Unique point Steam Type 5 CI For Steam processes between 121-135 °C



IT26-C

For Steam processes between 118-138 °C

